

Bid Evaluation Sheet For Haematology Analyzer 3Part

Part -I & II

Knock Down Criteria (Commercial Evaluation)

All evaluation parameters defined below are mandatory for compliance

Sr.No	Evaluation Parameters	M/S Meditec Instrument Co Karachi	M/S A.MI Sales & Services Karachi	M/S AMI Healthcare Faisalabad	M/S Mordin Dow Lahore	M/S Popular International Multan	M/S AH Traders Lahore	M/S Technostar Enterprises Lahore	M/S Human Healthcare Lahore
1	Complete Package / tender	YES	YES	YES	YES	YES	YES	YES	YES
2	Affidavit from Bidder	YES	YES	YES	YES	YES	YES	YES	NO
3	Bid Security	YES	YES	YES	YES	YES	YES	YES	YES
4	Bid Validity	YES	YES	YES	YES	YES	YES	YES	YES
5	Delivery Period	YES	YES	YES	YES	YES	YES	YES	YES
6	Exclusive Authorization/ Sole Agent Certificate by Manufacturer	YES	YES	YES	YES	YES	YES	NO	Expired
7	Technical & Engineering Capability (As defined for the specific tender in specifications)/ List	YES	YES	YES	YES	YES	YES	YES	YES
8	Certificate from manufacturer about the after sales services through agent or itself (in case specifically demanded in the specification)	YES	YES	YES	YES	YES	YES	NO	YES
9	Vendor past performance (Incase of unsatisfactory performance detail must be mentioned)	NO	YES	Not for Specific Product	Not for Specific Product	YES	Not for Specific Product	NO	YES
10	Availability of Relevant Tools and Testing/Calibration Equipment	YES	YES	YES	YES	YES	YES	YES	YES
11	Compliance of warranty as per tender	YES	YES	YES	YES	YES	YES	YES	YES
12	International Standard	China	YES	YES	YES	YES	NO	NO	NO
13	Imprinted product on Current official website	YES	YES	YES	YES	YES	YES	YES	YES






PART - III


KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Sl.No	SPECIFICATION COMPLIANCE/EVALUATION PARAMETERS										
1	Name of Equipment	Haematology Analyzer 3Part									
		Reseller	M/S Meditec Instrument Co Karachi	M/S A.M Sales & Services Karachi	M/S AM Healthcare Faisalabad	M/S Morfin Dow Lahore	M/S Popular International Multan	M/S AH Traders Lahore	M/S Technostar Enterprises Lahore	M/S Human Healthcare Lahore	
		Brand	Mindray	Nihon kohden	Sysmex	Swelab Boule Diagnostic	Acon	Bilimscl	Biobax	Vitanix D.OO	
		Model	BC-30	Celltac α+ MEK-1301	XP-100	Alfa Plus	HA-360	Br-Pw21	BK3200	Vitacount 3P	
		Country of Origin of Product/Model Number	China	Japan	Japan	Sweden	USA	Turkey	China	Solventia	
		Compliance with defined quality standards	No	Yes	Yes	Yes	Yes	No	No	No	
		Specification Compliance features wise:	No	No	No	No	Yes	No	No	No	
		Specifications:	1-Histogram for WBC,RBC not mentioned 2-No builtin QC	1-sample 60min/hr on open mode only 2-Histogram for WBC,RBC and platellet	1-No USB interference accessories 2-UPS online not mentioned	1-Not sine wave 2-No battery backup	Yes	1-No builtin QC	1-No Builtin QC 2- Autocalibrati on not mentioned 3-No accessories quoted	1-Histogram for WBC,RBC not specified	
		Technical Eligibility of Product:	No	No	No	No	Yes	No	No	No	
		Technical Eligibility of Firm:	No	Yes	No	No	Yes	No	No	No	
BID STATUS:	Not Responsive	Not Responsive	Not Responsive	Not Responsive	Responsive	Not Responsive	Not Responsive	Not Responsive			


Mrs. Sara Fatima
Bio Medical Engineer
D.G.Khan Medical College D.G.Khan


Dr. Zarnab Sattar
Senior Demonstrator
D.G.Khan Medical College D.G.Khan


Dr. Farooq Ahmad Malik
Associate Professor of Biochemistry
Chairman Technical Evaluation Committee
D.G.Khan Medical College D.G.Khan