

Technical Evaluation Sheet

D.G. Khan Medical College & Allied Institution D.G. Khan

Item Name

(INSTALLATION & COMMISSIONING OF MEDICAL GASES PIPELINE SYSTEM)

Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Part-I Commercial Parameters				
NTN & GST (Valid)	Yes	Yes	Yes	Yes
Bid Security	Yes	Yes	Yes	Yes
Registration with SECP/Registrar of Firms	Yes	Yes	Yes	Yes
Undertaking / Statement on Stamp Paper of Rs. 100/- that the Party including the director, and the owners is not a subject of bankruptcy proceedings, receivership, administration receivership, or any other form of liquidation	Yes	Yes	Yes	Yes
Average Annual Turnover as per tender requirement	Yes	Yes	Yes	Yes
The bidder shall comply the all the terms & conditions.	Yes	Yes	Yes	Yes
Affidavit from Bidder as per Form 8.6	Yes	Yes	Yes	Yes
Bid Validity	Yes	Yes	Yes	Yes
Delivery Period as per schedule of requirement	Yes	Yes	Yes	Yes
Remarks: Eligible / Not Eligible for further evaluations of PART-II)	Eligible	Eligible	Eligible	Eligible
Part-II Technical Parameters				
Valid Sole Agent Certificate for the Pakistan/Province in Punjab	Yes	Yes	Yes	Yes

Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Compliance with advertised specifications including Optional	Yes	Yes	Yes	Yes
Certificate from the manufacturer that the after sales	Yes	Yes	Yes	Yes
Compliance of quality standards for quoted product as mentioned in known down criteria	Yes	Yes	Yes	Yes
Brochures attached to verify the specifications	Yes	Yes	Yes	Yes
Foreign and Local trained technical staff as per tender	Yes	Yes	Yes	Yes
ISO Quality certification of the bidder	Yes	Yes	Yes	Yes
Workshop certificate and list of calibration and testing tools	Yes	Yes	Yes	Yes
Details of Spare Parts	Yes	Yes	Yes	Yes
Satisfactory local market business history and installed As mention in bidding Criteria. 8 Nos Purchase orders	Yes	No Business History of Brand CMS Medical Facilities and System Gmbh.	Not comply	Not comply
Remarks: Eligible / Not eligible for further evaluations of PART-III	Eligible	Not Eligible	Not Eligible	Not Eligible
Part-III Product Parameters				
Compliance with defined quality standards	Yes	Yes	Yes	Yes
Availability of quoted model on manufacturer website	Yes	Yes	Yes	Yes
AUTOMATIC MANIFOLD FOR OXYGEN.				
Brand	GENSTAR TECHNOLOGIES CO.,INC.USA	CMS Medical Facilities and System Gmbh.	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	GM3-TL-O2	GKE-3CD	YF-JYX1-SL-O2	Manifold-Auto-O2-3S
Country of Manufacturer	USA	Germany	China	UK

Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Country of Origin of Product	China	Germany	China	UK
Specification Compliance features wise	Yes	Not Comply Offered model not available on website and No Product performance history in the market	No VIE Connection.	Not Comply Offered model Capacity not mentioned in the BID
AUTOMATIC MANIFOLD FOR NITROUS OXIDE				
Brand	GENSTAR TECHNOLOGIES CO.,INC.USA	Greggersen Germany	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	GM2-DL-N2O	MC2050E	YF-JYX1-SL-N2O	Manifold-Auto-N2O
Country of Manufacturer	USA	Germany	China	UK
Country of Origin of Product	China	Germany	China	UK
Specification Compliance features wise	Yes	Yes	Yes	Not Comply Offered model Capacity not mentioned in the BID
GAS OUTLETS (BS/DIN SYSTEM)				
Brand	GENSTAR TECHNOLOGIES CO.,INC.USA	Greggersen Germany	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	38 Series	LEVANTE II & FORANO	P4-BSZD	TUWF-NG-G
Specification Compliance features wise	Yes	Yes	Yes	Not Comply Offered model AGSS not Venturi type
BED HEADS UNIT FOR ICU	Yes	Yes	Yes	Yes
BED HEAD UNIT FOR WARDS & PRIVATE ROOMS	Yes	Yes	Yes	Yes
ALARM AREA WITH REMOTE SENSORS				
Brand	GENSTAR TECHNOLOGIES CO.,INC.USA	Greggersen Germany	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK

Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Model	GM100M	Ventus EVO	YF-FMX	MOD-Compact-Alarm
Specification Compliance features wise	Yes	Yes	Not Comply offered not matched	Yes
Master Alarm				
Brand	Beacon Medaes Part of Atlas Copco	Greggersen Germany	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	MP125D	AEOLUS	YF-CYX	CPX-P-%
Country of Manufacturer	UK	Germany	China	UK
Country of Origin of Product	UK	Germany	China	UK
Specification Compliance features wise	Yes	Yes	Not Comply	Yes
MULTIPLE ZONE VALVE BOX				
Brand	GENSTAR TECHNOLOGIES CO.,INC.USA	Greggersen Germany	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	ZVB	Ventus EVO	YF-FMX	MOD-Compact-Alarm
Specification Compliance features wise	Yes	Yes	Yes	Yes
OXYGEN & AIR FLOW METER WITH S.S PROBE				
Brand	Flowmeter	Muller Medical	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK
Model	Easymed with TR200 for Adult and Neonate	Nanotube & MICROFLUX	YF-05D	FLOW-Q2-15-A
Country of Manufacturer	Italy	France	China	UK
Country of Origin of Product	Italy	France	China	UK
Specification Compliance features wise	Yes	Yes	Not Comply Neonate flowmeter no model offered and YF-05D not dial type	Not comply with Neonate dial type and no model offered
VACUUM REGULATOR WITH GAUGE / DIGITAL				
Brand	Flowmeter	Muller Medical	Yuyao Yufeng Medical Equipment Co. Ltd	Precision UK


Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Model	Easyvac 1000 plus MAK 1000 & Easyvac 250 plus MAK 1000	Nanovac1000 & Nanovac 250	YF-2X	HSUCK
Country of Manufacturer	Italy	France	China	UK
Country of Origin of Product	Italy	France	China	UK
Specification Compliance features wise	Yes	Yes	Not Comply model offered for Neonate	Not Comply No model offered for Neonate
COPPER PIPING				
Brand	Tonoli	SILMET	Tonoli	Precision UK
Model	FEIN	Not Quoted	Diff Size	CP
Country of Manufacturer	Italy	Italy	Italy	UK
Country of Origin of Product	Italy	Italy	Italy	UK
Specification Compliance features wise	Yes	Yes	Quoted model not verified from Catalogue	Yes
TRIPLEX MEDICAL VACUUM SYSTEM				
Brand	Beacon Medaes Part of Atlas Copco Group	No Detail	Connect Medical System	Precision UK
Model	MVAC2500T	HOSPIVAC G E200 R	0203-2106	MV-1350-T
Country of Manufacturer	UK	Not mentioned	UK	UK
Country of Origin of Product	China	Not mentioned	UK	UK
Specification Compliance features wise	Yes	Not Comply Tank Capacity 2500 L not confirmed	Yes	Not Comply with Tank capacity offered 1500L instead 2500L
MEDICAL COMPRESSED AIR STATION				
Brand	Beacon Medaes Part of Atlas Copco Group	No Detail Available	Connect Medical System	Precision UK
Model	Compressor G11 P MED, Dryer PMA-M1-10/7 & 2005704	No Detail Available	0102-4115	MA-1950-D
Country of Manufacturer	UK	No Detail Available	UK	UK
Country of Origin of Product	China	No Detail Available	UK	UK

Evaluation Parameters	1	2	3	4
	M/s	M/s	M/s	M/s
Name of Firms	Total Technologies (PVT) Ltd.	Radiant Medical	Multan Chemical Limited	Ghani Gases
Specification Compliance features wise	Yes	Not Comply	Yes	Yes
BID STATUS:	Responsive	Non Responsive	Non Responsive	Non Responsive


 Mrs Sara Fatima
 Bio Medical Engineer
 D.G. Khan Medical College
 Member


 Dr. Zamab Sattar
 Senior Demonstrator
 D.G. Khan Medical College
 Member


 Dr. Kashif Ali
 Assistant Professor (Medicine)
 D.G. Khan Medical College
 Member


 Dr. Syed Aushtar Abbas
 Associate Professor
 (Anesthesia)
 D.G. Khan Medical College
 Member


 Dr. Farooq Ahmad Malik
 Associate Professor (Community
 Medicine)
 D.G. Khan Medical College
 Member