



**OFFICE OF THE
PRINCIPAL D.G KHAN MEDICAL COLLEGE/ TEACHING
HOSPITAL DERA GHAZI KHAN
PH.NO-064-9260631 FAX NO.064-9260632**



**MINUTES OF THE MEETING OF GRIEVANCE COMMITTEE REGARDING
INSTALLATION & COMMISSIONING OF MEDICAL GASES PIPELINE SYSTEM
FOR THE SCHEME "ESTABLISHMENT OF EMERGENCY & OPD BLOCK AT
TEACHING HOSPITAL D.G.KHAN FOR THE YEAR 2025-26 HELD ON 29-04-2026
AT 11.00 AM**

The meeting of Grievance committee was held in the Committee room of D.G.Khan Medical College Dera Ghazi Khan on **29-04-2026 at 11.00AM** regarding the Grievances against Technical Evaluation report of Installation & Commissioning of Medical Gases Pipeline System for the Scheme "Establishment of Emergency & Opd Block at Teaching Hospital D.G.Khan for the Year 2025-26.

The following attended the meeting.

Sr.#	Name & Designation	
1	Dr.Saad Ullah Malik Professor of Ophthalmology D.G.Khan Medical College D.G.Khan	Chairman 
2	Dr.Uzma Shaheen Associate Professor of Pathology D.G.Khan Medical College D.G.Khan	Member 
3	Dr.Muhammad Haroon Bilal Assistant Professor of Medicine D.G.Khan Medical College D.G.Khan	Member
4	Dr. Haseeb Iqbal Senior Demonstrator D.G.Khan Medical College D.G.Khan	Member 
5	Mr.Muhammad Ismail Audit Officer D.G.Khan Medical College D.G.Khan	Member 

The meeting was started with the recitation from Holy Quran, The chairman welcomes all the participants and briefed them about proceeding of this meeting and then representatives of firms were called one by one for their grievances. All the Grievances were thoroughly discussed point to point and explained the reasons of being non-responsive/not eligible if any. The attendance of the representative of the firms was also recorded. The details of Grievances and decisions of Committee are as under:-

REGARDING INSTALLATION & COMMISSIONING OF MEDICAL GASES PIPELINE SYSTEM FOR THE SCHEME "ESTABLISHMENT OF EMERGENCY & OPD BLOCK AT TEACHING HOSPITAL D.G.KHAN FOR THE YEAR 2025-26 HELD ON 29-04-2026 AT 11.00 AM

Sr.No	Name of firm	Observation by technical committee	Reply of firm	Decision by the Grievance Redressel committee	Status
1.	M/S Ghani Chemical Industries Ltd	1-Satisfactory local market business history and installed as mention in bidding Criteria. 8 Nos Purchase orders(Not Comply) 2-AUTOMATIC MANIFOLD FOR OXYGEN(Specification Compliance features wise Not Comply Offered model Capacity not mentioned in the BID) 3-AUTOMATIC MANIFOLD FOR NITROUS OXIDE(Specification Compliance features wise Not Comply Offered model Capacity not mentioned in the BID) 4-OXYGEN & AIR FLOW METER WITH S.S PROBE(Specification Compliance features wise Not comply with Neonate dial type and no model offered) 5--VACUUM REGULATOR WITH GAUGE / DIGITAL (Specification Compliance features wise Not Comply o model offered for Neonate) 6-TRIPLEX MEDICAL VACUUM SYSTEM(Specification Compliance features wise Not Comply with Tank capacity offered 1500L instead 2500L)	The firm did not submit the documents.	The decision of technical evaluation committee is upheld	Not Responsive
2.	M/S Multan Chemical Ltd	1-Satisfactory local market business history and installed as mention in bidding Criteria. 8 Nos Purchase orders(Not Comply) 2-AUTOMATIC MANIFOLD FOR OXYGEN(Specification Compliance features wise No VIE Connection.) 3-OXYGEN & AIR FLOW METER WITH S.S PROBE(Specification Compliance features wise Not Comply Neonate flowmeter no model offered and YF-05D not dial type) 4-VACUUM REGULATOR WITH GAUGE / DIGITAL (Specification Compliance features wise Not Comply No model offered for Neonate) 5-copper piping (model not verified from catalog)	1-Our firm has provided comprehensive evidence of its market experience more than 21 performance certificate confirming satisfactory cognition. 2-we clarified that 3-we hereby undertake that the same model equipped with dial type flowmeters in full compliance with required specifications. 4-we clarified that the manufacturer provided a single unified model (YF-2X) suitable for both Adult and Neonate use. there both requirements were technically covered under one model specification. 5-we submitted that universally	The firm M/s Multan Chemicals do not comply with the bid document's clause r & s regarding Past Performance of the quoted product 8 references of manifolds, air plant and vacuum plant. • The quoted Manifold system of M/s Multan Chemicals do not have VIE Connection controlled by the manifold cabinet, and do not have complete automatic shifting with display of pressures of each source on manifold. • The offered Model of item Area Alarm is not available in TDS/brochure. • The offered Model of Item Flow Meter do not comply with the requirement for Neonate patients 0-5 lpm and the offered model is not dial type as required for	Not Responsive






3.

**M/S Radiant
Medical Pvt Ltd**

1-Satisfactory local market business history and installed as mention in bidding Criteria. 8 Nos Purchase orders(No Business History of Brand CMS Medical Facilities and System GmbH.)
2-AUTOMATIC MANIFOLD FOR OXYGEN(Specification Compliance features wise Not Comply Offered model not available on website and No Product performance history in the market)
3-TRIPLEX MEDICAL VACUUM SYSTEM(Specification Compliance features wise Not Comply Tank Capacity 2500 L not confirmed)
4-MEDICAL COMPRESSED AIR STATION (Specification Compliance features wise Not Comply)
Grievance from M/S Multan Chemical
We would like to bring to your attention the following concerns

1. Sole Distribution of CMS Medical Facilities and Systems GmbH
Ms Radiant Medical is not an authorize Distributor of CMS Medical Facilities and Systems GmbH, which raises serious concerns regarding the authenticity and compliance of their submitted bid.
2. Relevant Performance History
Furthermore, the firm appears to lack a verifiable performance record related to the CMS Medical Facilities and Systems GmbH, and the nature of the project. This absence of relevant experience may impact the quality, reliability, and successful execution of such a critical healthcare infrastructure project.

specified and identified by the size/dimensions rather than model numbers, which is standard industry practice.

1-we clarified that After Prebid meeting we have received amendments in which capacity of the O2 manifold was 120m3/h we have wrote letter copy enclosed the PVMS. dated: 03-01-2021 that specs received after prebid are not as per PVMS it is against the spirit of
In PVMS it is written (100-120m3/h) that was designed after discussion with all the interested vendors in the market with consensus so that maximum competition will be ensured institute has been given choice to select between 8090m3/h or 100-120m3.
Greggersen Germany which has 100 m3/h capacities as per PVMS.
Point.24 of bidding document clause 2.5.8 A. Technical evaluation criteria. Sub clause "r" under the heading of past performance: referred,
"The 90% of tender quantity of quoted product must have been installed in Pakistan"
It is submitted that referred item from CMs Germany is just one item and rest of the quoted brands are already installed in many institutes of Pakistan, order copies, installation reports, performance reports are enclosed with.


precise & accurate flow of oxygen for neonate patients.
• Model of Item Vacuum Regulator for Neonates is not offered / mentioned can low flow cannot be achieved for neonate patients.
So the decision of technical evaluation committee is upheld

1-the firm has failed to clarify the objection raised by the technical committee
2-the compressor is not quoted in technical bid
3-the firm has failed to provide the business history of oxygen manifold make CMs medical facilities and system The decision of technical evaluation committee is upheld

Not
Responsive

			So, we fulfill the 90% of tender quantity of quoted product must have been installed in Pakistan clause. Referred quoted item is from famous brand Cms medical Germany. 2-the offer model is on link https://www.cmsmed.de/medizinische-gasversorgung 3-that was missed in scanning due to human error and available in financial offer 4-that was missed in scanning due to human error and available in financial offer		
--	--	--	--	--	--

The meeting ended with the vote of thanks to and from the chairman.


Dr. Haseeb Iqbal
 Senior Demonstrator
 D.G.Khan Medical College D.G.Khan


Mr. Muhammad Ismail
 Audit Officer
 D.G.Khan Medical College D.G.Khan


Dr. Uzma Shaheen
 Associate Professor of Pathology
 D.G.Khan Medical College D.G.Khan

Dr. Muhammad Haroon Bilal
 Assistant Professor of Medicine
 D.G.Khan Medical College D.G.Khan


Dr. Saad Ullah Malik
 Professor of Ophthalmology
 D.G.Khan Medical College D.G.Khan
 (Chairman)